

EC Type-Approval Certificate

No. DK 0199.325

WTB...

NON-AUTOMATIC WEIGHING INSTRUMENT

Issued by DELTA Danish Electronics, Light & Acoustics

EU - Notified Body No. 0199

In accordance with the requirements for the non-automatic weighing instrument of EC Council Directive 2009/23/EC.

Issued to Bosche GmbH & Co. KG.

Reselager Rieden 3 49401 Damme GERMANY

In respect of Non-automatic weighing instrument designated WTB... with variants of

modules of load receptors, load cells and peripheral equipment.

Accuracy class III and IIII

Maximum capacity, Max: From 0.3 kg up to 999 999 kg

Verification scale interval: e = Max / n

Maximum number of verification scale intervals: $n \le 10000$ for single-interval and $n \le 3 \times 10000$ for multi-range and multi-interval (however, de-

pendent on environment and the composition of the modules).

Variants of modules and conditions for the composition of the modules are

Signatory: J. Hovgård

set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of OIML R76:2006, the European Standard EN 45501:1992/AC:1993 and WELMEC 2.1:2001.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 13 pages.

Issued on 2011-11-10 Valid until 2021-11-10 DELTA

Danish Electronics, Light & Acoustics

Venlighedsvej 4 2970 Hørsholm Denmark

Tel. (+45) 72 19 40 00 Fax (+45) 72 19 40 01 www.delta.dk VAT No. DK 12275110



Descriptive annex

	Contents	Page
1.	Name and type of instrument and modules	2
2.	Description of the construction and function	2
2.1	Construction	2
2.2	Functions	2
3.	Technical data	4
3.1	Indicator	4
3.2	Indicator - EEx version	5
3.3	Load receptors, load cells and load receptor supports	6
3.4	Composition of modules	7
3.5	Documents	7
4.	Interfaces and peripheral equipment	7
4.1	Interfaces	7
4.2	Peripheral equipment	8
5.	Approval conditions	8
5.1	Measurement functions other than non-automatic functions	8
5.2	Counting operation is not approved for NAWI	8
5.3	Compatibility of modules	8
6.	Special conditions for verification	8
6.1	Composition of modules	8
7.	Securing and location of seals and verification marks	8
7.1	Securing and sealing	8
7.2	Verification marks	9
8.	Location of CE mark of conformity and inscriptions	9
8.1	Indicator	9
9.	Pictures	11
10.	Composition of modules - example	13



1. Name and type of instrument and modules

The weighing instrument is designated WTB. It is a system of modules consisting of an electronic indicator connected to a separate load receptor and peripheral equipment such as printers or other devices, as appropriate. The instrument is a Class III or IIII, self-indicating weighing instrument with single-interval, multi-range or multi-interval.

The name of the instrument may have the designation "JOLLY" put in front and may be followed by alphanumeric characters for technical, legally or commercial characterisation of the instrument.

The indicators consist of analogue to digital conversion circuitry, microprocessor control circuitry, power supply, keyboard, non-volatile memory for storage of calibration and setup data, and a weight display contained within a single enclosure.

The modules appear from the sections 3.1, 3.2 and 3.3; the principle of the composition of the modules is set out in the sections 6.1 and 10.

2. Description of the construction and function

2.1 Construction

2.1.1 Indicator

The electronic indicator consists of two electronic boards: a main board bearing the microcontroller, the ADC and all other components, and a display board. The main board's microcontroller and analogue section are identical for all models.

The display has LED indication for: NET, zero, stable and weight unit (kg, g), and 6 digits with a height of 8 mm.

The enclosure is made of ABS plastics intended for mounting in a DIN rail with screw terminals in top and bottom for connection of power, load cell and interface cables.

Behind the cover glass of the front of the enclosure are 4 keys for operating the functions of the indicator.

All instrument calibration and metrological setup data are stored in the non-volatile memory.

The indicator is power supplied with 12 - 24 VDC.

2.1.2 Load receptors, load cells and load receptor supports

Set out in section 3.3.

2.1.3 Interfaces and peripheral equipment

Set out in section 4.

2.2 Functions

The weight indicating instruments are microcontroller based electronic weight indicators that require the external connection of strain gauge load cell(s). The weight information appears in the digital display located on the front panel and may be transmitted to peripheral equipment for recording, processing or display.

The primary functions provided are detailed below.



2.2.1 Display range

The weight indicators will display weight from –Max to Max (gross weight) within the limits of the display capacity.

2.2.2 Display test

A self-test routine is initiated at power up. The test routine turns on and off all of the display segments and indicators to verify that the display is fully functional.

2.2.3 Zero-setting

Pressing the "ZERO" key for 3 seconds causes a new zero reference to be established and ZERO annunciator to turn on indicating the display is at the centre of zero.

Semi-automatic zero-setting range: 4 % of Max. Automatic zero-tracking range: 4 % of Max. Initial zero-setting range: 20 % of Max.

Zero-setting is only possible when the load receptor is not in motion.

2.2.4 Zero-tracking

The indicators are equipped with a zero-tracking feature which operates over a range of 4 % of Max and only when the indicator is at zero (gross or net) and there is no motion in the weight display.

2.2.5 Tare

2.2.5.1 Semi-automatic tare

The instrument models are provided with a semi-automatic subtractive tare feature activated using the "TARE" key.

2.2.5.2 Preset tare

The indicators have a preset tare function for manually insertion of a tare value. The function can operate simultaneously with a semi-automatic tare, if the preset tare is activated first.

The value of the preset tare can be displayed temporary upon operator's request.

2.2.6 Printing

A printer may be connected to the optional serial data port. The weight indicator will transmit the current to the printer when the "PRINT" key is pressed.

The printing will not take place if the load receptor is not stable, if the gross weight is less than zero, or if the weight exceeds Max.

2.2.7 Alibi memory

The indicators are equipped with an alibi memory; however, the alibi memory must be enabled in the configuration of the indicator in order to be used.

When enabled pressing the "Print" key will store the displayed weight in alibi memory, if it is a stable legal weight not already stored.

The alibi memory is a circularly used buffer, but an alarm message is displayed each time the first record is overwritten.



2.2.8 Weighing unstable samples (optional)

The indicator has a function for weighing unstable samples. It can only be used, if it is enabled in the configuration.

2.2.9 Operator information messages

The weight indicator has a number of general and diagnostic messages which are described in detail in the user's guide.

2.2.10 Software version

The software version is displayed during the start-up of the indicator.

The version format is xx.yy.zz, where xx is the legal version no., while yy and zz are major and minor version numbers for changes and corrections not influencing the legal function of the software.

The tested software version is 01.09.15.

3. Technical data

The WTB weighing instrument is composed of separate modules which are set out as follows:

3.1 Indicator

The indicators have the following characteristics:

Type WTB...
Accuracy class III or IIII

Weighing range Single-interval or multi-interval (2 or 3)

Maximum number of verification scale intervals (n) 10000 for Class III

1000 for Class IIII

 $\label{eq:minimum input} \begin{tabular}{ll} Minimum input voltage per VSI & 0.2 μV \\ Maximum capacity of interval or range (Max_i): & n_i \times e_i \\ Verification scale interval, e_i = & Max_i/n_i \\ \end{tabular}$

Initial zero-setting range: \pm 10 % of Max Maximum tare effect: \pm 100 % of Max

Fractional factor (pi) 0.5
Excitation voltage 5 VDC

Circuit for remote sense Active, (see below)

Minimum input impedance 43 ohm
Maximum input impedance 1200 ohm

Connecting cable to load cell(s): See section 3.1.1

Supply voltage: 12 - 24 VDC, or 230 VAC Operating temperature range $\frac{\text{Min/Max} = -10 \text{ °C/} + 40 \text{ °C}}{\text{C}}$

Peripheral interface(s) See Section 4



3.1.1 Connecting cable between the indicator and the junction box for load cell(s), if any

3.1.1.1 4-wire system

Line 4 wires, shielded

Maximum length The certified length of the load cell cable, which shall be connect-

ed directly to the indicator.

3.1.1.2 6-wire system

Line 6 wires, screened

Option 1:

Maximum length $1315 \text{ m/mm}^2 \text{ (for n} = 10,000)$

Maximum resistance per wire 22.2 ohm

In case the (n) for the weighing instrument is less than (n) mentioned above, the following apply:

Option 2:

Coefficient of temperature of the span error of the indicator: Es = 0.001 [%/25K] Coefficient of resistance for the wires in the J-box cable: Sx = 0.0015 [%/ohm]

 $L/A_{max} = 295.86 / Sx * (emp/n - Es) [m/mm²] in which emp = p'i * mpe * 100/e$

From this, the maximum cable length for the weighing instrument may be calculated with regard to (n) for the actual configuration of the instrument.

Reference: See section 10.

The calculation program is obtainable by downloading at www.delta.dk/weighing.

3.2 Indicator - EEx version

The indicator has the following characteristics when connected through shunt-diode safety barrier type MTL7761ac and MTL7766Pac:

Type WTB...
Accuracy class III or IIII

Weighing range Single-interval or dual-interval

Maximum number of verification scale intervals (n) 7500 for Class III

1000 for Class IIII

 $\label{eq:minimum input} \begin{tabular}{ll} Minimum input voltage per VSI & 0.52 μV \\ Maximum capacity of interval or range (Max_i): & n_i \times e_i \\ Verification scale interval, e_i = & Max_i/n_i \\ \end{tabular}$

Initial zero-setting range: $\pm 10 \%$ of Max Maximum tare effect: $\pm 100 \%$ of Max

Fractional factor (pi) 0.5
Excitation voltage 5 VDC

Circuit for remote sense Active, (see below)



Minimum input impedance 43 ohm
Maximum input impedance 1200 ohm

Connecting cable to load cell(s): See section 3.1.1

Supply voltage: 12 - 24 VDC, or 230 VAC Operating temperature range $\frac{\text{Min/Max} = -10 \text{ °C/} + 40 \text{ °C}}{\text{C}}$

Shunt-diode safety barrier: MTL7761ac, MTL7766Pac, or similar

Peripheral interface(s) See Section 4

3.2.1 Connecting cable between the indicator and the junction box for load cell(s), if any

3.2.1.1 4-wire system

Line 4 wires, shielded

Maximum length The certified length of the load cell cable, which shall be connect-

ed directly to the indicator.

3.2.1.2 6-wire system

Line 6 wires, screened

Option 1:

Maximum length $1847 \text{ m/mm}^2 \text{ (for n} = 7,500)$

Maximum resistance per wire 31.2 ohm

In case the (n) for the weighing instrument is less than (n) mentioned above, the following apply:

Option 2:

Coefficient of temperature of the span error of the indicator: Es = 0.002 [%/25K] Coefficient of resistance for the wires in the J-box cable: Sx = 0.0012 [%/ohm]

 $L/A_{max} = 295.86 / Sx * (emp/n - Es) [m/mm²] in which emp = p'i * mpe * 100/e$

From this, the maximum cable length for the weighing instrument may be calculated with regard to (n) for the actual configuration of the instrument.

Reference: See section 10.

The calculation program is obtainable by downloading at www.delta.dk/weighing.

3.3 Load receptors, load cells and load receptor supports

Removable platforms shall be equipped with level indicators.

3.3.1 General acceptance of modules

Any load cell(s) may be used for instruments under this certificate of type approval provided the following conditions are met:

1) A test certificate (EN 45501) or OIML Certificate of Conformity (R60) respectively issued for the load cell by a Notified Body responsible for type examination under the Directive 2009/23/EC.



- 2) The certificate contains the load cell types and the necessary load cell data required for the manufacturer's declaration of compatibility of modules (WELMEC 2, Issue 5, 2009), and any particular installation requirements). A load cell marked NH is allowed only if humidity testing to EN 45501 has been conducted on this load cell.
- 3) The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in the above WELMEC 2 document, or the like, at the time of EC verification or declaration of EC conformity of type.
- 4) The load transmission must conform to one of the examples shown in the WELMEC 2.4 Guide for load cells.

3.3.2 Platforms, weigh bridge platforms

Construction in brief All-steel or steel-reinforced concrete construction, surface or pit

mounted

Reduction ratio 1

Junction box Mounted in or on the platform
Load cells Load cell according to section 3.2.1

Drawings Various

3.3.3 Bin, tank, hopper and non-standard systems

Construction in brief Load cell assemblies each consisting of a load cell stand assembly to

support one of the mounting feet bin, tank or hopper

Reduction ratio 1

Junction box Mounted on dead structure

Load cell according to section 3.2.1

Drawings Various

3.4 Composition of modules

In case of composition of modules, EN 45501 paragraph 3.5 and 4.12 shall be satisfied.

3.5 Documents

The documents filed at DELTA (reference No. A530871) are valid for the weighing instruments described here.

4. Interfaces and peripheral equipment

4.1 Interfaces

4.1.1 Load cell input

The connector pins for load cell connection are located on the bottom of the enclosure.

4.1.2 Other interfaces

The indicator may be equipped with one or more of the following protective interfaces,

- RS485, Profibus, DeviceNet and CANopen
- USB
- Digital output
- Digital input



Analogue output

The interfaces are characterised "Protective interfaces" according to paragraph 8.4 in the Directive and do not have to be secured.

4.2 Peripheral equipment

Connection between the indicator and peripheral equipment is allowed by screened cable.

The instrument may be connected to any simple peripheral device with a CE mark of conformity.

5. Approval conditions

5.1 Measurement functions other than non-automatic functions

Measurement functions that will enable the use of the instrument as an automatic weighing instrument are not covered by this type approval.

5.2 Counting operation is not approved for NAWI

The count shown as result of the counting function is not covered by this NAWI approval.

5.3 Compatibility of modules

In case of composition of modules, WELMEC 2 (Issue 5) 2009, paragraph 11 shall be satisfied.

6. Special conditions for verification

6.1 Composition of modules

The environmental conditions should be taken into consideration by the composition of modules for a complete weighing instrument, for example instruments with load receptors placed outdoors and having no special protection against the weather.

The composition of modules shall agree with section 5.3.

An example of a declaration of conformity document is shown in section 10.

7. Securing and location of seals and verification marks

7.1 Securing and sealing

Seals shall bear the verification mark of a notified body or alternative mark of the manufacturer according to ANNEX II, section 2.3 of the Directive 2009/23/EC.

7.1.1 Indicator

Access to the configuration and calibration facility requires that a calibration jumper is installed on the main board, or that the operator types first a password and the key looked up on a special key card delivered by the manufacturer.

The indicators have also an event counter, which increment each time the configuration is changed.

Sealing of the cover of the enclosure - to prevent access to the calibration jumper and to secure the electronics against dismantling/adjustment - is accomplished with a brittle plastic sticker. The sticker is placed so access to opening the enclosure is prohibited (see figure 4).



7.1.2 Indicator - load cell connector - load receptor

Securing of the indicator, load receptor and load cell combined is done in one of the following ways:

- Inserting the serial number of the load receptor as part of the principal inscriptions contained on the indicator identification label
- The load receptor bears the serial number of the indicator on its data plate.

7.1.3 Junction box for load cells

A junction box for load cells shall be sealed against opening with wire and seal or brittle plastic sticker(s).

7.1.4 Peripheral interfaces

All peripheral interfaces are "protective". Via the serial interface, zero and span adjust can be performed, similar to what the operator can, but it will increment the event counter as evidence for the infringement of the sealing. Apart from this, the interfaces neither allow manipulation with weighing data or legal setup, nor change of the performance of the weighing instrument in any way that would alter the legality of the weighing.

7.2 Verification marks

7.2.1 Indicator

A green M-sticker shall be placed next to the CE mark on the inscription plate.

The sticker with verification marks may be placed on or next to the inscription plate or on the front of the indicator.

7.2.2 Printers used for legal transactions

Printers covered by this type approval and other printers according to section 4.2, which have been subject to the conformity assessment procedure, shall not bear a separate green M-sticker in order to be used for legal transactions.

8. Location of CE mark of conformity and inscriptions

8.1 Indicator

8.1.1 **CE mark**

A sticker with the CE mark of conformity and year of production is located on the identification plate which is located on the enclosure of the weight indicator.

8.1.2 Inscriptions

Manufacturer's trademark and/or name and the type designation is located on the front panel overlay. Indelibly printed on a brittle plastic sticker located on a visible place near the front panel:

• Max, Min, e =

On the inscription plate:

• Manufacturer's name and/or logo, model no., serial no., type-approval certificate no., accuracy class, electrical data and other inscriptions.



Left to the manufacturer's choice as provided in section 7.1.2:

• Serial no. of the load receptor

8.1.2.1 Load receptors

On a data plate:

• Manufacturer's name, type, serial number, capacity

Left to the manufacturer's choice as provided in section 7.1.2:

• Serial no. of the indicator



9. Pictures



Figure 1 WTB indicator.



Figure 2 WTB_Devicenet indicator.





Figure 3 WTBPROFI indicator.



Figure 4 Sealing of WTB enclosure.



10. Composition of modules - example

COMPATIBILITY OF MODULES Ref.: WELMEC 2

Non-Automatic Weighing Instrument, single-interval Certificate of EU Type-Approval No: DK0199.325 **INDICATOR** WTB A/D (Module 1) Type: Accuracy class according to EN 45501 and OIML R76: Classind (I, II, III or IIII) Maximum number of verification scale intervals (I_{max}): 10000 nind Fraction of maximum permissible error (mpe) 0,5 p_1 [Vdc] Load cell excitation voltage: U_{exc} 5 [μV] [Ω] Minimum input-voltage per verification scale interval: 0,2 Δu_{min} Minimum load cell impedance: [%/25°C] Coefficient of temperature of the span error: Es 0,001 Coefficient of resistance for the wires in the J-box cable: Sx [%/Ω] 0,001 Specific J-box cable-Length to the junction box for load cells (L/A)_{max} [m / mm²] 6-wire (remote sense) Load cell interface: % of Max] Additive tare, if available: Initial zero setting range IZSR % of Max Temperature range T_{min}/T_{max} [°C] -10 40 Test report (TR), Test Certificate (TC) or OIML Certificate of Conformity LOAD RECEPTOR (Module 2) Type: Construction: Hopper! Fraction of mpe 0,5 1 Number of load cells: 3 Reduction ratio of the load transmitting device: $R=F_M/F_L$ Dead load of load receptor DL [% of Max] 35 Non uniform distribution of the load NUD 20 Correction factor: Q = 1 + (DL + T+ + IZSR+ + NUD) / 100 LOAD CELL ANALOG (Module 3) Type: Accuracy class according to OIML R60: Class_{LC} (A, B, C or D) C Maximum number of load cell intervals: 3000 nLC 0,7 Fraction of mpe: p₃ C Rated output (sensitivity): [mV/V] Input resistance of single load cell: R_{LC} $[\Omega]$ 409 $(v_{min\%} = 100 / Y)$ 0,008 Minimum load cell verification interval: [% of Emax] [kg] Rated capacity: 100 Minimum dead load, relative (E_{min /} E_{max}) * 100 0 [°c j**l** T_{min}/T_{ma} Temperature range 40 Test report (TR) or Test Certificate (TC/OIML) as appropriate **COMPLETE WEIGHING INSTRUMENT** Single-interva WTB w. hopper Manufacturer: Bosche Type: Class_{WI} (I, II, III or IIII) Accuracy class according to EN 45501 and OIML R76: Ш Fractions: $p_i = p_1^2 + p_2^2 + p_3^2$: 1.0 [kg] 150 Maximum capacity: Number of verification scale intervals: 3000 0.05 Verification scale interval [kg] Utilisation ratio of the load cell $\alpha = (Max / E_{max}) * (R / N)$ 0,50 Input voltage (from the load cels): $\Delta_{\rm u}$ = C * U_{exc} * α * 1000 / n 1,67 [uV/e] [mm² j l Cross-section of each wire in the J-box cable: 0.22 J-box cable-Length 25 [m] Temperature range to be marked on the instrument Not required T_{min}/T_{max} Peripheral Equipment subject to legal control <u>d no result</u> (WELMEC 2: 1) Class_{WI} Class_{ind} & Class_{LC} Class_{WI} (R76: 3.5.4.1) 1 - pi = pi 0,0 n_{max} for the class n_{max} for the class - n =(R76: 3.2) 7000 n (WELMEC 2: 4) 7000 n <= $n_{ind} - n =$ nind (R76: 4.12.2) <= 0 $n_{LC} - n =$ E_{min} DL*R/N (WELMEC 2: 6d) (DL * R / N) - E_{min} = 17.5 <= $v_{min} \star \sqrt{N} / R$ e - $(v_{min} * \sqrt{N / R}) =$ е (R76: 4.12.3) 0.036 Alternative solutions: or (if v_{min} is not given) (WELMEC 2: 7) e - ((E_{max}/ n_{LC}) * (\sqrt{N}/ R)) = $(E_{\text{max}} / n_{LC})_{\star} (\sqrt{N} / R)$ <= <= Δu (WELMEC 2: 8) 1,47 Δu_{min} $\Delta u - \Delta u_{min} =$ R_{Lmin} R_{LC} / N (WELMEC 2: 9) $(R_{LC}/N) - R_{Lmin} =$ 93 $(L / A)_{max}^{WI} - (L / A) =$ L/A (L / A)_{max} (WELMEC 2: 10) 4620 $(T_{max} - T_{min}) - T_{range} = E_{max} - (Q * Max * R / N) =$ <= (R76: 3.9.2.2) 20 Т, T_{max} - T_{min} * Max * R / N E_{max} (R76: 4.12.1) 17,5

Signature and date:

Conclusion PASSED

This is an authentic document made from the program "Compatibility of NAWI-modules version 3.2".

